

Certificate of Analysis

Synthetic Soil (Medium) OAS
Cat No. B2233– Certificate No. 437921

General

This Organic Analytical Standard (OAS) consists of a homogenous batch of Synthetic Soil Medium standard for use as a routine working microanalytical standard for the determination of Carbon, Nitrogen and Sulfur.

Certified Values and Uncertainty

The uncertainty in the certified value is expressed as expanded uncertainty, U, at 95% confidence and is calculated in accordance with ISO/IEC17025 according to GUM (Guidelines to Uncertainty in Measurement). Confidence limits include those due to sampling variation, weighing, calibration and measurement errors. The certified values are based upon the results of 30 determinations.

The certified value for Carbon, Nitrogen and Sulfur were determined by elemental analyser calibrated to Cystine 143d from National Institute of Standards and Technology (NIST), Maryland, USA.

	<i>Element</i>	<i>Certified Value</i> (% w/w)	<i>Uncertainty</i> (±%)
As bottled	Carbon	2.49	0.06
	Nitrogen	0.16	0.01
	Sulfur	0.037	0.009
	Total inorganic carbon	(0.15)	
	Total organic carbon	(2.34)	

Expiration of Certification

The certification of this OAS is valid until 10 January 2027 within the measurement uncertainties specified.

Storage and use

This OAS should be stored between 20°C to 25°C and should be kept tightly sealed away from light and moisture. It is non-hygroscopic under normal conditions and can be used without preliminary drying.

Certification Information

Bracketed figures are not certified values and are provided for information only.

The technical aspects involved in the preparation, certification, and issuance of this (In)Organic Analytical Standard (IAS/OAS) were carried out at Elemental Microanalysis Ltd, Okehampton, Devon EX20 1UB UK, Tel +44 1837 54446.

Email enquiries@microanalysis.co.uk.

For and on behalf of
Elemental Microanalysis Ltd

Jon Davies
Technical Director