

# Certificate of Analysis

Very Enriched Water  
 $^2\text{H}$  &  $^{18}\text{O}$  Water Standard - 30ml

Cat No. B2247 – Certificate no. 457387

## General

This standard is intended to provide a water sample of known isotopic composition that can be used to calibrate instrumentation and normalise unknown water samples. It is a homogenous batch of fresh distilled and deionised water to which enriched waters have been added.

It is not intended as a substitute for the primary reference materials distributed by the IAEA.

## Values and Uncertainty

The stable hydrogen and oxygen isotope values of this standard were determined by CF-IRMS using IAEA-607 and IAEA-608 for calibration. All samples were prepared by equilibration with hydrogen for  $\delta^2\text{H}$  and carbon dioxide for  $\delta^{18}\text{O}$ .

The results quoted are based on 28 determinations for  $\delta^2\text{H}$  and 64 determinations for  $\delta^{18}\text{O}$ . Uncertainty is expressed as one standard deviation ( $\sigma$ ) from the mean value.

<b>B2247</b>	<b><math>\delta^2\text{H}</math> v-SMOW (‰)</b>	<b><math>\delta^{18}\text{O}</math> v-SMOW (‰)</b>
<b>Very Enriched Water</b>	<b><math>2505.97 \pm 3.75</math></b>	<b><math>248.43 \pm 0.99</math></b>

## Storage

Once opened, this reference material should be stored carefully in a cool place with the lid done up tightly enough to avoid evaporation.

## Information

Calibrated and bottled September 2024.

The technical aspects involved in the production of this reference material were carried out by Elemental Lab, a division of Elemental Microanalysis. Okehampton, Devon, EX20 1UB, UK.  
Tel +44 1837 54446. <https://elementallab.co.uk/>

Email: [enquiries@microanalysis.co.uk](mailto:enquiries@microanalysis.co.uk)

For and on behalf of,  
Elemental Microanalysis Ltd.

David Castle  
R&D and Applications Specialist

