



Okehampton Business Park  
Exeter Road  
Okehampton  
Devon EX20 1UB  
Telephone: 01837 54446  
Fax: 01837 54544  
Web: [www.elementalmicroanalysis.com](http://www.elementalmicroanalysis.com)

**Certificate of Analysis**  
**Part No. B2345**  
**GREEN PETROLEUM COKE**  
**STANDARD**

Certificate Number 715901  
Page 1 of 1

**Dried Basis Value**

Mean Weight Percent Sulphur = 1.20  
Standard Deviation = 0.05

The certified mean values were statistically derived from the values reported through Round Robin.

**Method:**

ASTM D 4239

**Notes:**

Normal ASTM procedures should be employed when using this standard. This includes using the reproducibility and repeatability factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages.

**EXPIRATION DATE:**

THIS RM IS VALID FOR TWO YEARS FROM THE DATE OF OPENING

Elemental Microanalysis Limited