

Elemental Microanalysis Limited

Okehampton Business Park
Exeter Road
Okehampton
Devon EX20 1UB
Telephone 01837 54446/7
Fax 01837 54544

**Certificate of Analysis
Part No. B2372
Residual Oil Standard**

Certificate Number	191007
Date	17 April 2013
Page	1 of 1

**Weight Percent Sulphur = 2.500
Standard Deviation = 0.125**

The certified mean values were statistically derived from the values reported through Round Robin

Methods

ASTM D-1552
ASTM D-2622
ASTM D-5453-93

This certified reference material was traced to N.I.S.T. SRM 2717

Notes

Before use, the contents of the bottle should be mixed through vigorous shaking. The contents should not be exposed to air for a lengthy period of time.

Normal ASTM procedures should be employed when using this standard. This includes the reproducibility and repeatability factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

This is a Certified Reference Material (CRM), and is traceable to the above-mentioned standard. For good laboratory practice it is recommended that all standards be verified prior to use.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages.

This CRM is valid for two years from the date of purchase

Elemental Microanalysis Limited