



Okehampton Business Park
Exeter Road
Okehampton
Devon EX20 1UB
Telephone: 01837 54446
Fax: 01837 54544
Web: www.elementalmicroanalysis.com

Certificate of Analysis

Part No. B2386

SULFUR IN GREEN PETROLEUM COKE REFERENCE STANDARD

Certificate Number 211018

Page 1 of 1

Dried Basis Value

%Sulphur

Weight Percent Sulphur = 5.43

Expanded Uncertainty = ± 0.15

METHOD:

ASTM D 4239-17 AND ARI-035

This RM is traced to N.I.S.T. SRM 2718 and CANMET PC-2
Alpha- AR2721-700, AR2721-211201, AR756-561203, AR756-561117

The intended use of this reference material is for the calibration and quality validation of sulphur by resistance furnace, infra-red detection analysis as specified by ASTM 4239 or other valid test methods. The analytical values were derived by data points ($n=42$) and reported in mass fraction. The sample size used, and minimum sample size is approximately 300-500mg. The precision value represents the expanded degree of uncertainty at a 95% confidence level ($k=2$), utilizing ANOVA, ISO Guide 35, and the Guide to Uncertainty Measurement. Formal testing procedures should be followed when using this standard; this includes using the reproducibility and repeatability factors of the method for establishing overall method analytical uncertainty. When necessary, professional judgment is applied toward consideration of data and statistical information.

The material used in production of this reference standard was identified in accordance with ARI 041. The samples for round robin testing were selected in accordance with ARI 031. The above values relate only to the material used to produce this standard. The analytical samples were dried per the NMI used or corrected for moisture as per the test method.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages. The above values relate only to the material used to produce this standard. This certificate cannot be reproduced except in full. This bottle contains 50g, minus 60 mesh (250 micron) petroleum coke powder. While unable to determine a definite shelf life, when kept stored and sealed properly, this product should be reviewed 20 years from the date of certification. Once opened this certificate is valid for two years.

This Reference Material is traceable to the above-mentioned standards. For good laboratory practice, it is recommended that all standards be verified as fit for purpose prior to use.

EXPIRATION DATE:

THIS RM IS VALID FOR TWO YEARS FROM THE DATE OF OPENING

Certified April 23, 2019

Elemental Microanalysis Ltd