

Okehampton Business Park Exeter Road Okehampton Devon EX20 1UB Telephone: 01837 54446 Fax: 01837 54544

Web: www.elementalmicroanalysis.com

Certificate of Analysis
Part No. B2852
Sulphur in Crude Oil Standard

Certificate Number 941100 Page 1 of 1

Weight Percent Sulphur = 0.0102 Standard Deviation = 0.0009

> Methods: ASTM D 4294 ASTM D 2622

The above value represents the mean from the data, as reported by round robin.

NOTES:

Before use, the contents of the bottle should be mixed though vigorous shaking. The contents should not be exposed to air for a lengthy period.

Normal ASTM procedures should be employed when using this standard. This includes using the reproducibility and repeatability factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

This reference material (RM) should be verified prior to use against a primary standard material provided by a governing agency such as N.I.S.T. when available.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages.

EXPIRATION DATE: THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF OPENING

Elemental Microanalysis Ltd