

Okehampton Business Park Exeter Road Okehampton Devon EX20 1UB Telephone: 01837 54446 Fax: 01837 54544

Web: www.elementalmicroanalysis.com

Certificate of Analysis Part No. B2910 Lube Oil Standard

Certificate Number 831996 Page 1 of 1

Weight Percent Sulphur = 0.073 Standard Deviation = 0.002

The above value represents the mean from the data, as reported by round robin.

Methods: ASTM D 4294

This Reference Material was traced to N.I.S.T. SRM 1819 Oil I, II, III, IV, and V

NOTES:

Before use, the contents of the bottle should be mixed through vigorous shaking. The contents should not be exposed to air for a lengthy period of time.

Normal ASTM procedures should be employed when using this standard. This includes using the *reproducibility* and *repeatability* factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages.

EXPIRATION DATE: THIS RM IS VALID FOR TWO YEARS FROM THE DATE OF OPENING

Elemental Microanalysis Ltd