



Okehampton Business Park  
Exeter Road  
Okehampton  
Devon EX20 1UB  
Telephone: 01837 54446  
Fax: 01837 54544  
Web: [www.elementalmicroanalysis.com](http://www.elementalmicroanalysis.com)

**Certificate of Analysis**  
**Part No. B2917**  
**Lube Oil Standard**

Certificate Number 381016  
Page 1 of 1

**Weight Percent Sulphur = 0.0100**  
**Expanded Uncertainty = ± 0.0012**  
**(k=2, @95% confidence limit)**

**Method used for analysis: ASTM D 4294-16, and ARI-078**  
**NMI Standard(s) used: NIST SRM 1819a I, II, III**

The intended use of this standard is for the calibration and validation of sulphur in lube oil or petroleum-based products by XRF or other valid test methods. This standard was produced gravimetrically using high purity materials, with balances calibrated and checked by precision NIST traceable weights. The above ASTM test method and NMI references listed were used for testing. The uncertainty value represents the 95% confidence limit (k=2) derived from analysis (n=30), and possible errors in weighing. When necessary, professional judgment is applied toward consideration of data and statistical information. Various test procedures can be employed when using this standard. Refer to your instrument manufacturer or approved test method for specific applications. This includes using the reproducibility and repeatability factors for the test method you wish to employ. The sample size used for testing was placed into a removable sample cup, equipped with replaceable X-ray transparent plastic film and providing a sample depth of at least 4mm and a diameter of at least 10mm. The material used in production of this standard was sampled in accordance with ARI 041. The samples for round robin testing were selected in accordance with ARI 031. The above values relate only to the material used to produce this standard.

Before use, the contents of the bottle should be mixed by vigorous shaking. Any exposure to air and light should be kept to a minimum. Keep sealed and store upright under normal laboratory conditions. This bottle contains 100ml to be used as per your test method or instrument manufacturer. Unopened and stored properly this product has an indefinite shelf life. Once opened this certificate is only valid for two years.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event, shall Elemental Microanalysis Ltd be liable for incidental or consequential damages. This certificate cannot be reproduced except in full.

This Reference Material (working standard) is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

EXPIRATION DATE:  
THIS RM IS VALID FOR TWO YEARS FROM THE DATE OF OPENING

Certified October 27, 2016

Elemental Microanalysis Limited