

Okehampton Business Park Exeter Road Okehampton Devon EX20 1UB Telephone 01837 54446/7 Fax 01837 54544

Certificate of Analysis Part No. B2919 Lube Oil Standard

Certificate Number 8571099 Page 1 of 1

Weight Percent Sulphur = 5.010 Standard Deviation = 0.018

> Method ASTM D 4294

SRM 1819 OIL I, II, III, IV, AND V

The above value represents the mean from the data, as reported by round robin

Notes

Before use, the contents of the bottle should be mixed thoroughly by vigorous shaking of the bottle. The contents should not be exposed to the air for a lengthy period of time.

Normal ASTM procedures should be employed when using this standard. This includes using the reproducibility and repeatability factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages.

Expiration Date:
This CRM is valid for two years from the date of opening

Elemental Microanalysis Limited