

Analytical Results	
% Sulfur Mean = 0.50 Expanded Uncertainty = ± 0.03 (k=2, @95% confidence limit)	% Chlorine Mean = 1.00 Expanded Uncertainty = ± 0.05 (k=2, @95% confidence limit)
Method used for verification: EDXRF Scan	
This standard was produced gravimetrically using high purity materials, with balances calibrated and checked by precision NIST traceable weights.	

**The analytical results above are provided by an accredited reference material manufacturer with a current certification in ISO 17025 and 17034.*

The intended use of this Reference Material (RM) is for the calibration and or verification of sulfur and chlorine analysis in lube oil or similar materials by XRF or other valid testing methods.

The sample size used for testing was placed into a removable sample cup, equipped with replaceable X-ray transparent plastic film, and providing a sample depth of at least 4mm and a diameter of at least 10mm. Sample size and minimum sample size may be contingent upon your test method or instrumentation manufacturer recommendations.

The Period of Validity for this RM is 2 years after opening and should be reviewed 20 years after the date below. Any exposure to air and light should be kept to a minimum. Keep sealed and store upright under normal laboratory conditions.

This bottle contains 100ml of liquid to be used per the test method you follow. Before use, the contents of the bottle should be mixed by gentle mixing.

Refer to your test methods and or manufacturer manual for expanded uncertainties, repeatability/reproducibility factors.

For good laboratory practice, we recommend that all reference materials be verified as fit for purpose prior to use. Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd. be liable for incidental or consequential damages.

Certified on the 25th of May 2022.

Elemental Microanalysis Ltd