



Okehampton Business Park  
Exeter Road  
Okehampton  
Devon EX20 1UB  
Telephone 01837 54446/7  
Fax 01837 54544

**Certificate of Analysis**  
**Part No. B2944**  
**SULPHUR AND CHLORINE IN**  
**LUBE OIL STANDARD**

Certificate Number 507194  
Page 1 of 1

**% SULPHUR = LESS THAN 1 PPM**  
**% CHLORINE = LESS THAN 1 PPM**

**METHODS:**

1. ASTM D2622
2. ASTM D808

**Notes:**

Before use, the contents of the bottle should be mixed through vigorous shaking. The contents should not be exposed to air for a lengthy period.

Normal ASTM procedures should be employed when using this standard. This includes using the reproducibility and repeatability factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages.

**EXPIRATION DATE:**

**THIS CRM IS VALID FOR TWO YEARS FROM DATE OF OPENING**

Elemental Microanalysis Limited